



SomaScan[®] Assay v4.1

The SomaScan Platform is a powerful tool with applications in scientific research and health and wellness.

This technical note provides background on the platform itself, as well as performance metrics for the SomaScan Assay v4.1.

Introduction

The SomaScan Assay from SomaLogic is the only proteomic technology capable of measuring rapidly (high throughput), broadly (thousands of proteins simultaneously), and deeply (high- and low-abundance proteins).

The SomaScan Platform is a highly multiplexed, sensitive, quantitative, and reproducible proteomic tool for discovering previously unidentified biomarkers for drug discovery, pre-clinical and clinical drug development, and clinical diagnostics, across a wide range of important diseases and conditions.

The SomaScan Assay v4.1 measures approximately 7,000 unique human protein analytes in small volumes of biological samples. The assay was developed for performance in human serum and plasma and offers exceptional dynamic range, quantifying the relative levels of proteins in plasma that span 10 logs in abundance with excellent reproducibility.

The SomaScan Platform is enabled by the generation of protein-capture reagents called SOMAmer® (Slow Off-rate Modified Aptamer) reagents. SOMAmer reagents consist of short single-stranded DNA sequences that incorporate hydrophobic modifications, greatly expanding the physicochemical diversity of the large randomized nucleic acid libraries from which the SOMAmer reagents are selected.

The SomaScan Platform measures native proteins in complex matrices by transforming available binding sites on individual proteins into a corresponding SOMAmer reagent concentration, which is then quantified by hybridization to microarrays.

The assay takes advantage of SOMAmer reagents' dual nature as both protein affinity-binding reagents with defined three-dimensional structures and unique nucleotide sequences recognizable by specific DNA hybridization probes.

The SomaLogic laboratory is registered with the Centers for Medicare & Medicaid Services (CMS) under CLIA of 1988 and is accredited by the College of American Pathologists (CAP).

The SomaScan Platform is being applied to a wide range of diseases and conditions to deliver insights that enable biomarker discovery, diagnostics development, pharmaceutical discovery and development, and health management.

Previous versions of the SomaScan Assay have been applied successfully to biomarker discovery and validation in many pharmaceutical research and development projects, diagnostics discovery and development projects and academic research projects. A selected list of peer-reviewed articles can be found [here](#).

SOMAmer reagents are discovered using robust SELEX technology

SOMAmer reagents are single stranded DNA-based protein affinity reagents that benefit from aptamer technology developed over the past 30 years.^{1,2} The more recent proprietary innovation incorporates chemically modified nucleotides (Figure 1) that facilitate the binding to proteins, expanding the chemical diversity of standard aptamers and enhancing the specificity and affinity of protein-nucleic acid interactions.³




FIGURE 1 Modified nucleotides. Nucleotide triphosphate analogs modified at the 5-position (R) of uridine; 5-benzylaminocarbonyl-dU; 5-naphthylmethylaminocarbonyl-dU; 5-tryptaminocarbonyl-dU

These modified nucleotides are incorporated into nucleic acid libraries used for the iterative selection and amplification process called SELEX (Systematic Evolution of Ligands by EXponential enrichment) from which SOMAmer reagents are selected.⁴⁻⁶ Using a novel, proprietary SELEX process, SomaLogic has generated SOMAmer reagents to proteins that had been resistant to selection with unmodified nucleic acids (ACTG traditional aptamers).³

These chemical modifications in turn allow the SOMAmer reagent to tightly bind its target protein in ways never before possible⁷. A key advantage of this artificial selection process is that conditions can be tailored to select for desirable properties, including specificity, slow off-rate, and specific assay conditions.

SOMAmer reagents are selected against proteins in their native folded conformations and are therefore generally found to require an intact, tertiary protein structure for binding. As such, unfolded and denatured—and therefore presumably inactive—proteins are not detected by SOMAmer reagents. Crystal structures of SOMAmer reagents bound to their cognate protein targets indicate that the modified nucleotides contribute extensively to intramolecular contacts within the SOMAmer reagent and to intermolecular contacts with the protein targets.⁶

In Figure 2, the X-ray crystal structure of the IL-6 SOMAmer reagent bound to IL-6 demonstrates that the interactions between the SOMAmer reagent and its cognate protein are mainly mediated via the modified nucleotides. The co-crystal structures show very specific interactions between the SOMAmer reagent and target with binding site dimensions of 1100-1200 Å², similar to antibody–antigen interactions⁸.

The dissociation kinetics of a subset of SOMAmer reagents binding to their respective targets were determined using a solution-phase radiolabeled binding assay,³ and a subset were confirmed using the Biacore Flexchip surface plasmon resonance biosensor.

Biosensor results confirm slow dissociation off-rates, in the 10⁻⁵ to 10⁻⁴ s⁻¹ range, that correlate well with dissociation rate constants measured by solution-phase radiolabeled binding assays. Overall, SOMAmer reagents are analogous to high-quality antibodies that recognize intact tertiary protein structures. However, since they are made out of nucleic acids, SOMAmer reagents have several advantages over antibodies, such as tailored in vitro selection conditions, chemical synthesis, storage stability, and detection using sensitive and advanced DNA detection methods.




FIGURE 2 X-ray crystal structure of a SOMAmer reagent bound to IL-6. Modified nucleotides are shown in purple, and the DNA backbone and unmodified bases are in green.

SOMAmer Reagents to Approximately 7,000 Human Proteins

SOMAmer reagents have been created for human protein targets that cover a diverse set of biological processes, including cancer, inflammation and cardiovascular function, to name a few. Targets to date extensively cover major molecular functions including receptors, kinases, growth factors and hormones, and span a diverse collection of secreted, intracellular and extracellular proteins or domains.

SOMAmer Reagents are Chemically Synthesized, Stable and Rigorously Analyzed

After identification using SELEX technology, the SOMAmer reagents are chemically synthesized, purified and analyzed by methods including ultra high-performance liquid chromatography (UPLC), capillary gel electrophoresis (CGE) and mass

spectrometry (MS). Extensive functional analysis ensures consistent high performance of the SOMAmer reagents as quantitative affinity reagents. In order to test the specificity of SOMAmer reagents used in the SomaScan Assay for their respective initial target proteins, we perform a variety of characterization steps.

These steps include:

- In silico selection, procurement (when available), and direct SOMAmer reagent binding experiments in buffer with related proteins.
- Pull-down assays followed by MS-based and SDS gel-based analyses of the protein(s) bound by the SOMAmer reagent from biological matrices.

All types of affinity reagents (antibodies, traditional aptamers, etc.) are subject to specificity issues. Recognizing how critical the accuracy of the SomaScan Assay is for both research and clinical purposes, SomaLogic is committed to regular assessment of SOMAmer reagent specificity, and to transparency in communicating the results of those ongoing efforts.

Additional information can be found in the Characterization of SOMAmer Reagents Technical Note (SL00000430).

The Assay Steps

The SomaScan Assay quantitatively transforms the protein epitope availability in a biological sample into a specific SOMAmer-based DNA signal (Figure 3).

A SOMAmer-protein binding step is followed by a series of partitioning and wash steps to convert relative epitope concentrations into measurable nucleic acid signals that are quantified using DNA-hybridization microarrays.

Assay details are provided in the Appendix and in Gold et al., 2010.³ The readout in relative fluorescent units (RFU) is directly proportional to the amount of target epitope in the initial sample.

Achieving the 10¹⁰ Dynamic Range: SOMAmer Reagent Mixes

The large dynamic range of the SomaScan Assay in plasma and serum results from the optimization of SOMAmer reagent measurements across three serial dilutions of the sample (Figure 4). A specific SOMAmer reagent is only present in one of the three dilution groups.

The least concentrated sample is designed to detect the most abundant proteins, and the most concentrated solution is designed to detect the least abundant proteins. Based on reported literature values, the SomaScan Assay robustly measures analytes spanning a range of 10 logs in human plasma or serum (e.g., albumin to interleukins or interferons).

SomaScan Assay Characterization

The SomaScan Assay v4.1 has been validated in human EDTA plasma and serum. The SomaScan Assay has excellent reproducibility. In plasma, half of the SOMAmer reagents demonstrated a median Coefficient of Variation (CV) less than 5.3% and only 10% of the SOMAmer reagents demonstrated a median CV of 10.7% or higher. All values were determined in the multiplex assay, profiling approximately 7,000 reagents simultaneously.

By measuring thousands of proteins at a time in each blood sample with the SomaScan Platform, it is possible to uncover a precise set of specific protein

changes that provide information on current status and future trajectory for virtually every disease or condition of interest. For example, we can find repeatable patterns of protein changes that are associated with clinical indications (e.g., cardiovascular events) or fitness attributes (e.g., VO₂ Max).

Each set of specific protein patterns, in turn, becomes the basis of a specific SomaSignal Test that can be ordered by researchers who desire clinical assessments of study participants. The current suite of available RUO SomaSignal Tests can be found on the [SomaLogic website](#).

SomaLogic Quality Systems

The SomaScan Assay is performed under the SomaLogic Quality System (QS) in a laboratory that follows CLIA standards for a laboratory developed test.

The assay is performed in a facility that contains both access and environmental control. Equipment within the facility is maintained, calibrated, and operated in compliance with controlling Standard Operating Procedures (SOPs).


Equipment and associated software are validated for their intended uses in support of the SomaScan Assay.

Method validation has been completed for processes that could impact the performance of the SomaScan Assay. SOPs cover the incoming receipt, inspection, and release of raw materials to ensure that the materials used in the production of assay reagents or directly in the assay maintain the performance requirements established during the development of the SomaScan Assay.

FIGURE 3 SomaScan Assay.


STEP 01

SOMAmer reagents (purple) are synthesized with a fluorophore, photocleavable linker, and biotin.




STEP 02

SOMAmer reagents bound to streptavidin beads are used to capture proteins from a complex mixture of proteins (tan).




STEP 03

Unbound proteins are washed away, and bound proteins are tagged with biotin.




STEP 04

UV light breaks the photocleavable linker, releasing complexes back into solution.




STEP 05

Non-specific complexes dissociate while specific complexes remain bound.




STEP 06

A polyanionic competitor (green) prevents rebinding of non-specific complexes.



STEP 07

Biotinylated proteins (and bound SOMAmer reagents) are captured on new streptavidin beads.



STEP 08

SOMAmer reagents are released from the complexes by denaturing the proteins. Fluorophores are measured after hybridization to complementary sequences on a microarray chip. The fluorescence intensity detected on the microarray is related to the amount of available epitope in the original sample.








FIGURE 4 SomaScan Assay dynamic range. SOMAmer reagent mixes are prepared to achieve optimal detection in mixtures with a large range of concentrations. Shown here is the dilution distribution of SOMAmer reagents for SomaScan Assay v4.1 with plasma and serum.

Typical Protein Levels	fM - pM	nM	µM
 Plasma Dilution	 1:5	 1:200	 1:20,000
	+	+	+
Subset of SOMAmer Reagents	~85%	~12%	~3%

Summary

The SomaScan Platform is a powerful, highly multiplexed platform for discovering novel biomarkers during drug discovery, pre-clinical and clinical drug development, and for the development of clinical diagnostics, across a wide range of clinically important diseases.

Table 1 summarizes the SomaScan Assay v4.1 metrics in plasma. SomaLogic's SomaScan Platform technology provides significant advantages in sample size, cost, time, multiplexing capability, measurement range, and flexibility of readout over many alternate protein biomarker platforms.

Metric	Condition	Results
Sensitivity in Buffer	Median LOD (from representative subset)	125 fM or 5.3 pg/mL
Measurement Range	Span of proteins measured in plasma or serum	10 ¹⁰
Precision	Median Total % CV in plasma	5%
Sample Volume	Human serum or plasma (per sample)	130 µL
Multiplex Size	Total number of measurements	7,596
	SOMAmer reagent measurements	7,288

TABLE 1 Summary of SomaScan v4.1 Assay metrics to human targets.

References

1. Ellington AD & Szostak JW. (1990). In vitro selection of RNA molecules that bind specific ligands. *Nature* 346:818-22.
2. Tuerk C & Gold L. (1990). Systematic evolution of ligands by exponential enrichment: RNA ligands to bacteriophage T4 DNA polymerase. *Science* 249:505-10.
3. Gold L, et al. (2010). Aptamer-based multiplexed proteomic technology for biomarker discovery. *PLoS One* 5(12): e15004.
4. Vaught JD, et al. (2010). Expanding the chemistry of DNA for in vitro selection. *J. Am. Chem. Soc.* 132:4141-51.
5. Eaton BE. (1997). The joys of in vitro selection: chemically dressing oligonucleotides to satiate protein targets. *Curr. Opin. Chem. Biol.* 1:10-6.
6. Davies DR, et al. (2012). Unique motifs and hydrophobic interactions shape the binding of modified DNA ligands to protein targets. *Proc. Natl. Acad. Sci. USA* 109:19971-6.
7. Rohloff J, et al. (2014). Nucleic acid ligands with protein-like side chains: Modified aptamers and their use as diagnostic and therapeutic agents. *Mol. Ther. Nucleic Acids* 3: e201.
8. Ramaraj T, et al. (2012). Antigen-antibody interface properties: Composition, residue interactions, and features of 53 non-redundant structures. *Biochim. Biophys. Acta.* 1824:530-2.

Appendix

Details of the SomaScan Assay

The first step of the SomaScan Assay is the dilution of a biological sample of interest. The sample dilutions are incubated with the respective SOMAmer reagent mixes that have been attached to streptavidin (SA)-coated beads.

The beads are washed to remove non-specifically associated proteins and other matrix constituents. Proteins that remain bound to SOMAmer reagents are tagged using an NHS-biotin reagent. SOMAmer complexes and unbound SOMAmer reagents are released from the SA beads using ultraviolet light that cleaves a photo-cleavable linker within the SOMAmer reagent construct into a solution containing an anionic competitor.

Non-specific interactions dissociate, and the anionic competitor solution prevents them from reforming while specific complexes are maintained³. The photo-cleavage eluate, which contains all SOMAmer reagents (some bound to a biotin-labeled protein and some free), is separated from the beads and then incubated with a second streptavidin coated bead that binds the biotin-labeled proteins and the biotin-labeled protein-SOMAmer complexes.

The free SOMAmer reagents are removed during subsequent washing steps. In the final elution step, protein-bound SOMAmer reagents are released from their proteins using denaturing conditions and recovered. These SOMAmer reagents are then quantified by hybridization to custom DNA microarrays. The cyanine-3 signal from the SOMAmer reagent is detected on microarrays.

Data method: [Experimental Controls for Data Standardization and Quality Control Processing](#)

The SomaScan Assay is performed using 96-well plates; eleven wells are allocated for control samples used to control for batch effects and to estimate the accuracy, precision, and buffer background levels of the assay. Five pooled Calibrator Control replicates, three pooled Quality Control (QC) replicates, and three buffer (no protein) replicates are run on each plate.

For core sample types, Calibrator and QC replicates are created by pooling samples of the same type from presumed healthy donors. Twelve Hybridization Control SOMAmer reagents not exposed to sample proteins are added during the SOMAmer reagent elution step to control for readout variability.

The control samples are run repeatedly during assay qualification and robust point estimates are generated and stored as references for each SOMAmer reagent result for the Calibrator and QC samples. The results are to be used as references throughout the life of the SomaScan Assay version.

Data Standardization

Raw SomaScan Assay data may contain systematic biases from many sources, such as technical variation introduced by the readout, pipetting errors, or consumable reagent changes; more significantly, pre-analytical variance related to sample collection methods or inherent sample variation in overall protein levels leads to additional nuisance variance. Data standardization procedures are used to mitigate technical variation.

Data standardization is comprised of normalization and calibration which are routine numerical procedures developed to remove systematic biases in raw assay data after microarray feature aggregation. In general, normalization is a sample-by-sample adjustment to overall signals within dilution bins

(plasma and serum) or signaling bins (urine). Calibration is an overall plate adjustment and a SOMAmer-by-SOMAmer adjustment that decreases between-plate variability.

Each normalization method computes a scale factor, or set of scale factors, for each sample or SOMAmer reagent that is subsequently applied to the signal on the appropriate results in the report.

The data standardization steps include the following:

- Hybridization normalization
- Intra plate signal normalization of Calibrator and Buffer (no protein) replicates

- Plate scale standardization and Calibration using a global calibrator reference
- Signal normalization of the QC replicates using a global signal normalization reference
- QC check of the median of QC replicate values to the global QC reference standard specific for the pooled QC lot on the plate
- Signal normalization of the individual samples using a global signal normalization reference

For more details on the data standardization process see document our Data Standardization and File Specification Technical Note (SL00000442).



SL00000572 Rev 4: 2023-03

SomaScan Assay v4.1

SomaLogic® SomaScan® SOMAmer® and associated logo are trademarks of SomaLogic Operating Co., and any third party trademarks used herein are the property of their respective owners

© 2023 SomaLogic, Inc. | 2945 Wilderness Pl. Boulder, CO 80301 | Ph 303 625 9000 | www.somallogic.com